



Luminescence free counting of ^3H
facilitated by
Hidex 300 SL/600 SL TDCR triple
coincidence counters

Risto Juvonen

LSC 2017 Copenhagen, 4.5.2017



Hidex 300 SL and 600 SL Automatic TDCR Liquid Scintillation Counters

- Automatic vial counters with triple-PMT detector facilitating:
 - Exceptionally high counting efficiency,
 - Absolute activity counting of beta isotopes (H-3, C-14,..) and Cherenkov isotopes (Y-90, Sr-89,..) using TDCR method without external radioactive standard source
 - Luminescence free counting mode
 - a/b separation with 2D graphical calibration & validation tool



Luminescence interference

Luminescence is one of the main interferences in scintillation counting

Single photon event increasing the background and increases results variation

Methods to reduce and correcting luminescence are

- (1) Dark adaptation of sample
- (2) Chemical methods
- (3) Temperature control
- (4) Counting region settings
- (5) Delayed coincidence counting

Many labs are counting samples also with repeats, and removing outliers

- How to find outlier?
- Do I have to recount the sample or can I calculate the results using the data of remaining repeats?

Luminescence Free Counting method (LF method) for ^3H in water

Method is based on counting region (ROI) settings

With double coincidence counters the conventional method based on counting regions (ROIs) works relatively well for higher energy isotopes like ^{14}C but yields in too low counting efficiency for ^3H water samples.


Improved LF method was tested for ^3H water samples and requires triple coincidence counter like Hidex 300 SL or 600 SL.

HIDEX


Triple detector array
and coincidences

R S T: Triple Coincidence Counts

R S: Double Coincidence Counts
R T: Double Coincidence Counts
S T: Double Coincidence Counts



Luminescence do not interfere with triple counts



In Luminescence Free mode (LF mode) only Luminescence Free Counts are recorded:

- 1) All Triple Coincidence counts
- 2) Pure Double Coincidence counts at above luminescence ROI (> ch 120)

Improved eff. Allows also ^3H water measurements in LF mode

Performance comparison of Lum Free mode and Normal count mode for ^3H in water

Hidex 300 SL 425-020, 8 + 12 ml sample, 10 x 50 min, $k_a = k_b = 1.645$

	Normal Mode	LF Mode
Bkg	4.7 CPM	2.8 CPM
Eff.	37 %	24 %
Ld	2.6 Bq/L 1.2 DPM	3.1 Bq/L 1.5 DPM
Outliers	High risk	Low risk

IAEA TRIC2012 ³H Reference samples in LF mode

Sample	Hidex 300 SL, LF mode				TRIC Ref. labs		TRIC Top lab	
	8 mL + 12 mL UGuLLT, 20 x 50 min						2000 min	
	Deuterium		Spike					
	TU (² H)	Unc.	TU (spike)	Unc.	TU (known)	Unc.	TU (known)	Unc.
TRIC-20	0	0	0	0	0	0,10	0,02	0,01
TRIC-21	0,26	0,16	0,28	0,17	0,43	0,10	0,44	0,02
TRIC-22	1,13	0,18	1,17	0,18	1,12	0,10	1,12	0,15
TRIC-23	2,71	0,20	2,81	0,20	2,74	0,07	2,70	0,04
TRIC-24	4,25	0,11	4,45	0,21	4,37	0,13	4,45	0,06
TRIC-25	7,57	0,29	7,61	0,25	7,51	0,23	7,45	0,09

*Data provided by IAEA Hydrogeology lab, Vienna

Hidex 300 SL can be used to produce accurate data on LF mode for low level ³H water samples.

Quench Correction in LF mode

TDCR cannot be used directly in LF mode as luminescence interferes with double coincidence counts (c_d)

$$\mathbf{TDCR} = \frac{c_t}{c_d + c_t} = \frac{c_t}{c_{all}}$$

Available methods:

1. Constant quench correction factor (preferred when quenching is constant)
2. External std curve method
3. Correction of TDCR with Chemi CPM

$$\mathbf{TDCR} = \frac{c_t}{c_d - \mathit{Chemi} + c_t} = \frac{c_t}{c_{all} - \mathit{Chemi}}$$

4. Triple Coincidence Channel Ratio method
 - Ratio of Triple Coincidences in two different region used as quench parameter
=> (tROI1/tROI2) vs. Eff.

Summary

- Luminescence do not interfere with triple coincidence counts
- LF mode = all triples + pure doubles above luminescence ROI
⇒ Improved eff. allows also ^3H water measurements in LF mode
- In LF mode the Eff. and Bkg are reduced, Ld only slightly higher
- Total counting time is reduced as in LF mode counting can be started more or less immediately after sample preparation
- Low risk for outliers
- In LF mode the preferred quench correction method is the use of Constant Quench Correction Factor (other methods are applicable)
- Hidex 300 SL and 600 SL provide accurate data in LF mode for low level ^3H water samples



Thank You!

Risto Juvonen
Product manager
risto.juvonen@hidex.com

Lemminkäisenkatu 62
FI-20750 Turku

info@hidex.com
www.hidex.com

Tel. +358 10 843 5570